



KENYA REVENUE
AUTHORITY

ISO 9001:2015 CERTIFIED

ADDENDUM/ CLARIFICATION "SET 1"

21st January 2022.

To all Prospective bidders,

REF: KRA/HQS/NCB-024/2021-2022: SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF VARIOUS TESTING EQUIPMENTS FOR KRA LABORATORY.

Kenya Revenue Authority wishes to inform prospective bidders of the clarifications / addendum highlighted below:

NO.	Query reference from Tender	Bidders Questions	KRA's Response
1	LOT 1: Supply, Delivery, Installation and Commissioning of One (1) Unit Kjeldahl Analysis System at KESRA Mombasa.	a) Kindly specify if you need a digester with or without lift. b) Please indicate if you require a system complete with a titrator, or a system which would be upgradable to include a titrator later or a basic distillation unit. c) You have requested for Spare and service parts and accessories necessary for and the application of the system. kindly note that these accessories come as standard with the equipment. Please clarify if you need them additionally quoted.	Prospective bidders are advised that: A) The requirement is for Supply, Delivery, Installation and Commissioning of One (1) Unit Kjeldahl Analysis System fully functional complete with digester and Titrator. B) Digester with Lift, distillation and Titrator units and their accessories is a requirement. Bidders to comply with this requirement. C) the accessories come as standard or package with the equipment. Therefore, item number 1.7 in Lot 1 specifications has been expunged. Service and parts shall be covered by Annual Maintenance Costs as provided in the Revised Price Schedule for Lot 1 attached below as Annex 1.


[Signature]
20/01/22



KENYA REVENUE
AUTHORITY

ISO 9001:2015 CERTIFIED

2	Lot 2: Supply, Delivery, Installation and Commissioning of One (1) Unit Density Meter at KESRA Mombasa. 11-1-2022	a) It is indicated that spare and service parts and accessories necessary for operation and application of the requirement are required. Kindly specify the duration for which the spares offer should cover to allow bidders to quote for the required quantities.	a) the accessories come as standard or package with the equipment. Therefore, item number 18 in Lot 2 specifications has been expunged. Prospective bidders are required to use the attached Revised Price Schedule in Annex 1 that contains Annual Maintenance Cost (AMC) for Lot 2 when responding to the requirement.
3	Lot 2: Supply, Delivery, Installation and Commissioning of One (1) Unit Density Meter at KESRA Mombasa.	a) You have indicated that for controls, touchscreen, optional keyboard, mouse, barcode reader and gesture control are required. Please note that these are optional accessories, and the equipment can work without them. In this line, please confirm whether we should quote for them optionally or not. b) Also, you have indicated that you require a computer system with windows 10 and printer as well as the software for the PC control. Please note that the equipment is able to operate as a standalone without PC Control, that is, can be controlled via the touchscreen display. Please confirm whether the optional PC control via software is necessary and we should quote for it optionally or whether we can supply the standalone equipment.	a) The controls, touchscreen, optional keyboard, mouse, barcode reader and gesture control have been expunged . Bidders are therefore advised <u>NOT</u> to quote for the same. b) The PC control via Software is a requirement. Bidders are therefore advised to comply with this requirement.
4	Lot 3: Supply, Delivery, Installation and Commissioning of One (1) Unit of Drying Oven for Precise Drying of Samples to Facilitate Chemical Analysis at KESRA Mombasa. 10-1-2022	In the specifications, you have indicated that calibration should be at three (3) free selectable temperature values. Kindly specify if it is required and select your preferred temperature value(s) for the same.	Item 1.13 (Three (3) free selectable temperature values) in Lot 3 Specifications has been expunged and replaced with Calibration at + 160 °C . Prospective bidders to comply with this new requirement.


20/10/2022



KENYA REVENUE AUTHORITY

ISO 9001:2015 CERTIFIED

5	Lot 5: Supply, Delivery, Installation and Commissioning of One (1) Unit UV-Vis Spectrophotometer at KESRA Mombasa. 10-1-2022	You have indicated you require Performance Validation tests suitable for EP and USP mandated testing. Please note this is applicable only to the pharmaceutical industry. You have also requested for 21CFR compliance software also a requirement to the Pharma. Kindly confirm if both are needed as they come at an additional cost.	EP and USP performance has been expunged. 21CFR compliance software is not a requirement and has been expunged. Bidders are therefore required NOT to quote for the above two items.
6	LOT 8: Supply, Delivery, Installation and Commissioning of one (1) Refractometer at KESRA Mombasa. 11-1-2022	In the scope description you have stated a refractometer apparatus is used for measuring refractive index of gases, liquids, and translucent solids. Please note that it is not applicable to measure gases. Kindly specify the required application for the same for us to offer the appropriate refractometer.	The Refractometers is for liquid and solid phases only. The gas element has been expunged. Bidders are therefore advised to comply.
7	LOT 8: Supply, Delivery, Installation and Commissioning of one (1) Refractometer at KESRA Mombasa.	Also kindly specify whether you require a portable or benchtop type refractometer. Kindly specify the refractometer application and whether you require a portable or benchtop type.	Benchtop refractometer is a requirement. Bidders are advised to comply. The refractometer application is for liquid and solid phases only. A Benchtop refractometer is a requirement. Prospective bidders are therefore advised to comply.
	Price schedules for: <ul style="list-style-type: none">• Lot 1,• Lot 2,• Lot 5• Lot 8		Prospective bidders are hereby advised that the price schedules for Lot 1, Lot 2, Lot 5 and Lot 8 have been <u>expunged</u> and replaced with: Revised Price Schedules attached as Annex 1.

[Handwritten signature]
22/10/2022

Note:

·Bidders are advised to:

Acknowledge receipt of this addendum published and uploaded on the KRA Website on **21st January, 2022.**

The clarification / addendum form part of the bidding document and is binding to the bidder. All other terms and conditions of the tender remain the same. You are therefore required to immediately acknowledge the receipt of this addendum.

Regards,



Benson Kiruja

For: Deputy Commissioner - Supply Chain Management
Js



KENYA REVENUE
AUTHORITY

ISO 9001:2015 CERTIFIED

ANNEX 1: REVISED PRICE SCHEDULES FOR LOT 1, LOT 2, LOT 5 AND LOT 8

For Lot 1: Supply, Delivery, Installation and Commissioning of One (1) Unit Kjeldahl Analysis System

No	Item Description	Qty	Unit Price in Kshs (Inclusive of applicable taxes)	Total Cost in Kshs (Inclusive of applicable taxes)
1	Lot 1: Supply, Delivery, Installation and Commissioning of One (1) Unit Kjeldahl Analysis System	1		
	Annual Maintenance Cost (AMC) for Year 2	1		
	Annual Maintenance Cost (AMC) for Year 3	1		
	TOTAL COST IN KSHS FOR LOT 1 TO BE CARRIED TO THE FORM OF TENDER			

Note: In year 1, the equipment shall be under warranty. AMC will therefore not be applicable.


21/1/22



KENYA REVENUE
AUTHORITY

ISO 9001:2015 CERTIFIED

For Lot 2: Supply, Delivery, Installation and Commissioning of One (1) Unit Density Meter

No	Item Description	Qty	Unit Price in Kshs (Inclusive of applicable taxes)	Total Cost in Kshs (Inclusive of applicable taxes)
1	Lot 2: Supply, Delivery, Installation and Commissioning of One (1) Unit Density Meter	1		
	Annual Maintenance Cost (AMC) for Year 2	1		
	Annual Maintenance Cost (AMC) for Year 3	1		
TOTAL COST IN KSHS FOR LOT 2 TO BE CARRIED TO THE FORM OF TENDER				

Note: In year 1, the equipment shall be under warranty. AMC will therefore not be applicable.

For Lot 5: Supply, Delivery, Installation and Commissioning of One (1) Unit UV-Vis Spectrophotometer

No	Item Description	Qty	Unit Price in Kshs (Inclusive of applicable taxes)	Total Cost in Kshs (Inclusive of applicable taxes)
1	Lot 5: Supply, Delivery, Installation and Commissioning of One (1) Unit UV-Vis Spectrophotometer	1		
	Annual Maintenance Cost (AMC) for Year 2	1		
	Annual Maintenance Cost (AMC) for Year 3	1		
TOTAL COST IN KSHS FOR LOT 5 TO BE CARRIED TO THE FORM OF TENDER				


2011/12



KENYA REVENUE
AUTHORITY

ISO 9001:2015 CERTIFIED

For Lot 8: Supply, Delivery, Installation and Commissioning of one (1) Refractometer

No	Item Description	Qty	Unit Price in Kshs (Inclusive of applicable taxes)	Total Cost in Kshs (Inclusive of applicable taxes)
1	Lot 8: Supply, Delivery, Installation and Commissioning of one (1) Refractometer	1		
	Annual Maintenance Cost (AMC) for Year 2	1		
	Annual Maintenance Cost (AMC) for Year 3	1		
TOTAL COST IN KSHS FOR LOT 8 TO BE CARRIED TO THE FORM OF TENDER				

Note: In year 1, the equipment shall be under warranty. AMC will therefore not be applicable.

21/1/22